

# Declaration of Conformity

**Manufacturer:** Harvard Apparatus, a Division of Harvard Bioscience Inc<sup>1</sup>  
**Address:** 84 October Hill Road, Holliston, Massachusetts 01746-1388, USA

We herewith declare that the following products:

**Product Name:** TC202A Bipolar/Monopolar Temperature Controller  
**Models:** TC202A (65-0045)  
TC202AM (65-0046)

**Are in conformity of the following applicable European regulations and directives:**

2014/35/EU Low Voltage directive (LVD)  
2014/30/EU Electromagnetic Compatibility directive (EMC)  
2012/19/EU Waste electrical and electronic equipment directive (WEEE)<sup>2</sup>

**Standards used to demonstrate conformity include:**

EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use, General Requirements  
EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use –EMC requirements<sup>3</sup>

<sup>1</sup>Harvard Bioscience holds an ISO 9001:2008 certificate BSI FM 629138 for its Quality Management System

<sup>2</sup>Classified as belonging to equipment categories 8 or 9

<sup>3</sup>Classified and tested as class B equipment in accordance with CISPR 11 definition in a basic electromagnetic environment. This equipment has also been tested and found to comply with the limits for a class A digital device, pursuant to CFR Title 47 part 15 of the FCC rules.

I, the undersigned, hereby declare that the equipment specified is in conformity with the relevant harmonised Union legislation. Signed for and on behalf of Harvard Apparatus at Holliston, USA



Signed:

A handwritten signature in black ink, appearing to read 'Mark Davis'.

Date:

August 10, 2016

**Mark Davis**  
Harvard Bioscience, Senior Director of  
Engineering